



Portable Gas Monitor
GX-3R
Data Logger Management Program
SW-GX-3R(EX)
Operating Manual

RIKEN KEIKI Co.,Ltd.

2-7-6 Azusawa, Itabashi-ku, Tokyo, 174-8744, Japan
Website: <http://www.rikenkeiki.co.jp/>

1. Introduction

The operating procedures and precautions described in this operating manual apply only for use in accordance with the stipulated purposes. Riken Keiki rejects all liability in cases involving use of the program in ways not described in this manual.

This operating manual omits descriptions of basic operations like command selection and dialog box settings for Microsoft Windows 7, Windows 8, and Windows 10. If you are using Windows for the first time, read the Windows manual and familiarize yourself with basic Windows operations before proceeding.

WARNING

The CD on which this program is provided is a CD-ROM.
Do not attempt to play this CD on a regular audio CD player.
High audio volumes may damage your ears or speakers.

CAUTION

Requires pointing device.

This software requires the use of a pointing device such as a mouse or touchpad.
It cannot be used with a keyboard alone.

1-1. Software purpose and features

This software program is designed to import data collected using the data logger function of the GX-3R into a PC for various purposes.

Importing data collected using the data logger function into a PC offers the following benefits:

- Allows collected data to be listed.
- Allows collected data details to be displayed in graph or table form.
- Allows graph and table data to be printed and stored as hard copies.
- Retains records of past data.
- Eliminates the need to write down data on paper by hand.
- Helps pinpoint which devices need calibration.
- Simplifies the management of multiple devices.

NOTE

- Unauthorized reprinting or reproduction of the contents of this document in part or in whole is strictly prohibited, unless otherwise stipulated by law.
- The contents of this document are subject to change without notice to allow product improvements.
- The user must accept the separately provided Software License Agreement before using this product. Opening the product package constitutes consent to the terms of the license agreement.
- Use this software solely for the intended purpose. Modifying the product or using the product in a manner departing from this manual may affect product safety and quality. Riken Keiki rejects all liability for any incidents or damages arising from failure to observe the warnings indicated above.

(c) Copyright 2018 Riken Keiki Co., Ltd. All rights reserved.
Riken Keiki Co., Ltd. owns the copyright to this software.

Microsoft Windows 7, Microsoft Windows 8, and Microsoft Windows 10 are registered trademarks of Microsoft Corporation in the US and other countries.

Contents

	Page
1. Introduction.....	2
1-1. Software purpose and features.....	2
2. Installing and Uninstalling.....	4
2-1. Operating environment precautions.....	4
CAUTION: Precautions regarding handling of the CD-ROM.....	4
2-2. Installing the software.....	4
2-3. Installation procedure.....	5
CAUTION: Saving past data before reinstalling.....	8
CAUTION: Precautions for installing on Microsoft Windows 7, Microsoft Windows 8, and Microsoft Windows 10.....	8
2-4. Uninstallation procedure.....	9
3. Operating procedures.....	11
3-1. Download screen.....	11
① Receiving data from the GX-3R.....	12
② Receiving device information data.....	13
③ Downloading data.....	14
④ Downloading all main unit data.....	15
⑤ Clearing GX-3R main unit data.....	16
⑥ Turning off the power for the GX-3R main unit.....	16
⑦ Switching to automatic processing.....	17
3-2. Instrument Information screen.....	18
① Data source type.....	19
② Status information.....	19
③ Calibration history information.....	20
④ Sensor alarm setpoint information.....	20
3-3. Data screen.....	21
① Deleting data.....	22
② Viewing data details.....	22
③ Summary display area.....	23
3-4. Data View screen.....	26
① Selecting table or graph.....	27
② Sending to printer.....	28
CAUTION: Printer setup precautions.....	30
③ Saving to file.....	31
④ Viewing data summary at the same time.....	32
⑤ Table details.....	33
⑥ Graph details.....	34
3-5. Last Calibration screen.....	35
① Selecting display details.....	36
② Sending to printer.....	37
③ Deleting data.....	38
④ Changing the password.....	39
3-6. Set screen.....	40
① Setting font and graph colors.....	41
② Changing main unit status.....	43
③ Updating changes in GX-3R main unit.....	44
4. Data Maintenance.....	45
4-1. Data storage configuration details.....	45
4-2. Backing up.....	45
5. Usage Precautions.....	46
6. Troubleshooting.....	47
7. IrDA Specifications.....	48
7-1. Infrared communication.....	48
8. File Organization.....	49
8-1. Current directory immediately after installation.....	49
8-2. Current directory during operation.....	49
9. Software Function Specifications.....	50

2. Installing and Uninstalling

2-1. Operating environment precautions

This program is compatible with Microsoft Windows 7, Microsoft Windows 8, and Microsoft Windows 10. The program is not compatible with other operating systems.

This program requires up to approximately 40 MB of free hard disk space to install. It may require additional space, depending on the number of data samples. Make sure sufficient disk space is available.

CAUTION

Precautions regarding handling of the CD-ROM
<ol style="list-style-type: none">1. CD-ROM storage Do not store in locations subject to direct sunlight or high temperatures and humidity.2. CD-ROM drive type Do not insert in slot-loading CD-ROM drives. The label on the CD-ROM may prevent the CD-ROM from ejecting properly. Load the CD-ROM into a tray-loading CD-ROM drive.

2-2. Installing the software

Insert the install CD containing this program into the CD-ROM drive of your PC. The installation screen will appear automatically after a short while.

Do the following if the PC does not support automatic CD-ROM startup:

1. Open the CD-ROM drive in Explorer.
2. Double-click on the file "setup.exe".

2-3. Installation procedure

● Launch setup

The following screen will appear after the CD-ROM is inserted or setup.exe is launched.



Click the "Next" button.

● Accept license agreement

The following screen will appear:




To install the software, click the "Next" button. To abort the process, click the "Cancel" button.

CAUTION: Make sure you have read and fully understand the terms of the software license agreement before installing the software.


● User information

Click the “Next” button to display the following screen:



The screenshot shows a window titled "GX-3R - InstallShield Wizard". The main heading is "Customer Information" with a sub-heading "Please enter your information:". There are two text input fields: "User Name:" containing the text "test" and "Organization:". At the bottom, there are three buttons: "< Back", "Next >" (highlighted in blue), and "Cancel". The "InstallShield" logo is visible in the bottom left corner.

Click the “Next” button.

● Destination folder

The screenshot shows a window titled "GX-3R - InstallShield Wizard". The main heading is "Destination Folder" with a sub-heading "Click Next to install to this folder, or click Change to install to a different folder.". Below this, there is a folder icon and the text "Install GX-3R to: C:\Program Files\GX-3R\". A "Change..." button is located to the right of the folder path. At the bottom, there are three buttons: "< Back", "Next >" (highlighted in blue), and "Cancel". The "InstallShield" logo is visible in the bottom left corner.

Click the “Next” button.

- **Start setup**



Click the "Install" button to begin installing.

The following screen will appear once setup is complete:

- **Complete**



The program can be used as soon as setup is complete.

CAUTION**Saving past data before reinstalling**

Note the following precautions if reinstalling the program:

1. Uninstall the program before reinstalling.
2. If the program is uninstalled after it has been used, certain files will remain undeleted. One such file is the "GX3R.mdb" database file. If you wish to save past data, save this file to another location before deleting the folder.

CAUTION**Precautions for installing on Microsoft Windows 7, Microsoft Windows 8, and Microsoft Windows 10**

This software requires libraries for various drivers for the Windows system. Installing the software automatically initiates the processing required to incorporate these libraries.

However, if you are using Microsoft Windows 7, Microsoft Windows 8, or Microsoft Windows 10, you will be asked to install system libraries with administrator privileges.

Follow the instructions displayed to log in as an administrator and install the system libraries.

You will then be asked to restart Windows. When Windows restarts, log in once again as a general user, then install the application.

(Installing the libraries and application involves simply launching setup.exe on the CD-ROM. Administrator privileges are required only if the necessary libraries are not present in the system folder.)

Using with a network connection

Check the following before installing on a PC running Microsoft Windows 7, Microsoft Windows 8, or Microsoft Windows 10 and connected to a network.

IrDA communication uses certain TCP/IP-based technology (communication technology used in the Internet, etc.) and uses a special communication group and IP address.

For this reason, communications may be blocked in certain cases by strong Internet security software.

Where possible, run the software on a PC not connected to a network.


If use on a PC connected to a network is unavoidable, carefully assess the security settings before use.

2-4. Uninstallation procedure

● Startup

To uninstall the software, click “Start” on the taskbar, click “Settings”, and launch the Control Panel.


Control Panel (example)



Double-click to open “Programs and Features” in the Control Panel.

- **Select GX-3R**

Double-clicking "Programs and Features" opens the following window:



- **Start deletion**

Select (right-click) "GX-3R", then click the "Uninstall" menu. Or select (click) "GX-3R", then click the "Uninstall" button.

Click "Yes" to begin uninstalling.


CAUTION: The message "Do you want to remove the shared file?" may appear when uninstalling. Select "No"; selecting "Yes" may affect other applications.

3. Operating procedures

Start the program by clicking the "GX-3R" shortcut on the desktop or click the Start menu and start the "Program".

3-1. Download screen

The Download screen follows the splash screen.



To start data communication, place the GX-3R in a suitable location, start this program, and power on the main unit. The program will automatically determine whether data communication is possible; if so, it will enter reception standby mode.

① Receiving data from the GX-3R

● Main unit preparation


1. Start this software.
2. Place the GX-3R in a location where communication is possible with the power turned off.
3. Turn on the power for the GX-3R.

CAUTION:


Be sure to place the GX-3R in a location where communication is possible before turning on the power.

Connection cannot be made if moved to a location where communication is possible with the power already on.

The LCD on the GX-3R will appear as shown on the right. →
(This may be somewhat difficult to read due to the display limitations of the GX-3R LCD.)




The information area displays the following information:



CAUTION:

If the details shown in the information area differ from those shown here, power off the GX-3R main unit and check and change the location of the main unit as necessary before turning on the power once again.

The "Status" area changes to indicate communication is possible.



Serial Port:

- Communication possible: Green The PC port is available for use.
- Communication not possible: Red


Connect:

- Standby: Gray
- Communicating: Green

② Receiving device information data

● Receive device information

First, click the "Instrument Information" button to receive device information data.



An animated display will appear while data is being received.

Once the "Instrument Information" data has been received, the details shown in the "Instrument" area will be updated and the "Download Logger Data" and "Clear Logger Data" buttons enabled.


A screenshot of the instrument information interface. It features a 'Download Commands' section with five buttons: 'Complete Download', 'Instrument Information', 'Download Logger Data', 'Clear Logger Data', and 'Power Off'. Above these buttons is an 'Instrument' section with three fields: 'Serial No. (20 Characters)', 'Station ID (16 Characters)', and 'User ID (16 Characters)'. The 'Station ID' field contains 'S ID 001' and the 'User ID' field contains 'U ID 001'. A callout box points to these fields with the text 'Updated to show details of the GX-3R main unit'. Another callout box points to the 'Download Logger Data' and 'Clear Logger Data' buttons with the text 'The "Download Logger Data" and "Clear Logger Data" buttons are enabled.'

③ Downloading data

- Trend data
- Event data

After clicking the “Instrument Information” button and downloading the device information data, the “Download Logger Data” button is enabled.

Click the “Download Logger Data” button.



The information area displays the data receiving status.

CAUTION:


Since access to other data is not permitted, other download buttons and the “Set” button are disabled while data is being downloaded.

④ Downloading all main unit data

● All data

Clicking the “Complete Download” button downloads all data, including “Instrument Information”, “Interval Trend”, “Alarm Trend”, “Alarm Events”, “Trouble Events”, and “Power ON/OFF Events” data.

Click the “Complete Download” button.



The information area displays the data receiving status.

CAUTION:

Since access to other data is not permitted, other download buttons and the “Set” button are disabled while data is being downloaded.

⑤ Clearing GX-3R main unit data

● Clear data

Clicking the “Clear Logger Data” button clears all data from the GX-3R.

Click the “Clear Logger Data” button.



Click “Yes” to begin clearing the data.

⑥ Turning off the power for the GX-3R main unit

● Power off

Clicking the “Power Off” button turns off the power for the GX-3R main unit and resets the PC serial port.

1. Click the “Power Off” button.



Click “Yes” to begin the process of powering off the GX-3R main unit and to reset the PC serial port before switching to standby to await data from the main unit.



⑦ Switching to automatic processing

● Automatic download mode

1. Select the “Automatic Download” checkbox (if not already selected).



Click “Yes” to switch to automatic processing.
Click “No” to cancel the mode change.


● Automatic removal after download

All data is downloaded automatically by the PC when the GX-3R main unit is powered on. The GX-3R main unit will then power off.

Data cannot be downloaded manually while automatic processing is underway.

Automatic processing can be configured to automatically delete downloaded data after it is downloaded.

1. Select the “Automatic Data Removal” checkbox.



Click “Yes” to automatically delete all data inside the GX-3R after it is downloaded.

* This is a convenient way to reduce download times when repeating the Download → Delete → Download procedure several times.

3-2. Instrument Information screen

Click the “Instrument Information” button on the right-hand side of the screen to display the following screen. This screen lists device information data about the connected GX-3R main unit.

The screenshot shows the 'Instrument Information [Connected]' window. It contains several sections: 'GX-3R Status' with input fields for Serial No., Station ID, and User ID; 'Calibration History' table; 'Last Bump Test' table; and 'Warning and Alarm point' table. A sidebar on the right has buttons for Download, Instrument Information, Data, Last Calibration, Set, and Exit. Callouts point to specific data points and the Instrument Information button.

Calibration History

Gas	Calib.Date	Before	After	A.Cal.	Cal.Due(Days)
CH4(100%LEL)	1/1/2018	0	0	50	Now
O2(40.0%)	1/1/2018	0.0	0.0	12.0	Now
H2S(200.0ppm)	1/1/2018	0.0	0.0	25.0	Now
CO(2000ppm)	1/1/2018	0	0	50	Now

Last Bump Test

Gas	Bump Test Date	Test Result	Concentration	mp Test Due(Da
CH4(100%LEL)	1/1/2018	0	0	Now
O2(40.0%)	1/1/2018	0.0	0.0	Now
H2S(200.0ppm)	1/1/2018	0.0	0.0	Now
CO(2000ppm)	1/1/2018	0	0	Now

Warning and Alarm point


Gas	Warning	Alarm	AlarmH	STEL	TWA
CH4(100%LEL)	10	50	50	----	----
O2(40.0%)	18.0	18.0	25.0	----	----
H2S(200.0ppm)	1.0	10.0	10.0	5.0	1.0
CO(2000ppm)	25	50	50	200	25

CAUTION: This screen is read-only. Data cannot be edited on this screen. → See 3-6. Set screen.
Data is not displayed if the “Instrument Information” data has not been downloaded.

① Data source type

● Data source information


The indication will be “[Connected]” if information about the connected multi-gas monitor main unit is displayed.



② Status information

● Information details

Displays the “Serial No.,” “Station ID,” and “User ID” stored inside the main unit. CAUTION: These boxes are read-only and cannot be edited.



③ Calibration history information

● Calibration history details

Calibration History					
Gas	Calib.Date	Before	After	A.Cal.	Cal.Due(Days)
CH4(100%LEL)	1/1/2018	0	0	50	Now
O2(40.0%)	1/1/2018	0.0	0.0	12.0	Now
H2S(200.0ppm)	1/1/2018	0.0	0.0	25.0	Now
CO(2000ppm)	1/1/2018	0	0	50	Now

Details:

Gas: Measured gas names (full-scale units)
 Calib.Date: Date of last calibration
 Before: Concentration before last calibration
 After: Concentration/calibration failure after last calibration
 A.Cal: Automatic calibration concentration
 Cal.Due (Days): Warranty period for uncalibrated state (A warning is displayed in red 1 month before calibration expires.)

● Bump test history details

Last Bump Test				
Gas	Bump Test Date	Test Result	Concentration	mp Test Due(Da
CH4(100%LEL)	1/1/2018	0	0	Now
O2(40.0%)	1/1/2018	0.0	0.0	Now
H2S(200.0ppm)	1/1/2018	0.0	0.0	Now
CO(2000ppm)	1/1/2018	0	0	Now

Details:

Gas: Measured gas names (full-scale units)
 Bump Test Date: Date of last bump test
 Test Result: Concentration result for last bump test
 Concentration: Calibration gas concentration for last bump test
 Bump Test Due (Days): Warranty period for state with no bump test (A warning is displayed in red 1 month before the bump test expires.)

④ Sensor alarm setpoint information

● Details


Warning and Alarm point					
Gas	Warning	Alarm	AlarmH	STEL	TWA
CH4(100%LEL)	10	50	50	----	----
O2(40.0%)	18.0	18.0	25.0	----	----
H2S(200.0ppm)	1.0	10.0	10.0	5.0	1.0
CO(2000ppm)	25	50	50	200	25

Details:

Gas: Measured gas names
 Warning: 1st alarm setpoint concentration
 Alarm: 2nd alarm setpoint concentration
 AlarmH: 3rd alarm setpoint concentration
 STEL: STEL alarm setpoint concentration
 TWA: TWA alarm setpoint concentration

3-3. Data screen

Click the “Data” button on the right-hand side of the screen to display the following screen. This screen lists the downloaded data.



This screen can be used in the same way as Windows Explorer. However, the following operations are not available:

1. Renaming data
2. Moving data to other locations

The Explorer-style folders are displayed hierarchically in order of serial number, station ID, and user ID.

The folder and data names have the following formats:

Folder name: 03/11 = Data for March 2011

File name: 22111930_3EB = Interval trend for 11:19:30 on 22nd (date and time of logging start)


al26150419_3EB = Alarm trend for 15:04:19 on 26th (date and time of alarm occurrence)

The number of data samples allowed in each folder is limited by the PC's hard disk capacity. To maintain acceptable response times, you should back up data files each year. 4. See 4. Data Maintenance.

① Deleting data

● Delete


1. Click and select the data (folder) to delete.
2. Right-click without moving the mouse.



Click "Delete" on the "Delete" menu that appears.


● Password input

1. Enter the password at the prompt, then click the "Continue" button.



CAUTION: Clicking "Continue" without entering a password will cancel the delete.

2. The following message will appear when you enter the correct password and click the "Continue" button.



Click the "Yes" button to delete the data.
Click the "No" button to cancel data deletion.

② Viewing data details

● To data details

1. Click the data the details of which you wish to view. Confirm that the summary appears in the summary display area, then click the "View Data" button.
Or:
 2. Double-click the data the details of which you wish to view.
- For information on how to operate the Data View screen: → Refer to 3-4. **Data View screen.**

③ Summary display area

● Details

A summary of the data is displayed if the data selected is normal data.
Interval trend

Property	Value			
Name	iv15140007_-----			
Sampling Date	6/15/2018 2:00:07 PM to 6/15/2018 2:04:49 PM			
Serial No.	-----			
Station ID	S_ID_001			
User ID	U_ID_001			
Data Count	36			
Interval Time (sec)	10			
Gas(FullScale) - base:	CH4(100%LEL)	O2(40.0%)	H2S(200.0ppm)	CO(2000ppm)
Avg	1 %LEL	*****	0.0 ppm	0 ppm
Max	51 %LEL	*****	0.0 ppm	0 ppm
Max Date/Time	06/15 14:00:07	*****	06/15 14:00:07	06/15 14:00:07
Min	*****	*****	*****	*****
Min Date/Time	*****	*****	*****	*****

- Name: Data name
- Sampling Date: Date and time of measurement start and end
- Serial No./Station ID/User ID: GX-3R main unit status
- Data Count: Number of data samples
- Interval Time (sec): Sampling interval (seconds)
- Gas (FullScale) - base: Gas (full-scale) - base
- Avg: Gas average value
- Max: Gas data maximum value
- Max Date/Time: Date and time of maximum value detection
- Min: Gas data minimum value
- Min Date/Time: Date and time of minimum value detection
- Warning: 1st alarm setpoint
- Alarm: 2nd alarm setpoint
- AlarmH: 3rd alarm setpoint
- STEL: STEL alarm setpoint
- TWA: TWA alarm setpoint

Alarm events

DateTime	Gas	Event
5/30/2018 4:17:23 PM	CH4(100%LEL)	WARNING
5/30/2018 4:17:23 PM	CH4(100%LEL)	ALARM
5/30/2018 4:17:23 PM	CH4(100%LEL)	ALARM H

- DateTime: Date and time of event occurrence
- Gas: Gas generated
- Event: Event type

Alarm trend

Property	Value
Name	al15140008_-----
Alarm Date/Tir	6/15/2018 2:00:08 PM
Serial No.	-----
Station ID	S_ID_001
User ID	U_ID_001
Data Count	720
Interval Time (5
Gas(FullScale	13OH(100%LE O2(40.0%) 12S(200.0ppm CO(2000pp
Value	51 %LEL ***** 0.0 ppm 0 ppm
Warning	10 %LEL 18.0 % 1.0 ppm 25 ppm
Alarm	50 %LEL 18.0 % 10.0 ppm 50 ppm
AlarmH	50 %LEL 25.0 % 10.0 ppm 50 ppm
STEL	***** ***** 5.0 ppm 200 ppm

- Name: Data name
- Alarm Date/Time: Date and time of alarm occurrence
- Serial No./Station ID/User ID: GX-3R main unit status
- Data Count: Number of data samples
- Interval Time (sec): Sampling interval
- Gas (FullScale) - base: Gas (full-scale) - base
- Value: Concentration at time of alarm occurrence
- Warning: 1st alarm setpoint
- Alarm: 2nd alarm setpoint
- AlarmH: 3rd alarm setpoint
- STEL: STEL alarm setpoint
- TWA: TWA alarm setpoint

Calibration history

DateTime	Gas	Before	After
6/14/2018 11:49:24 AM	CH4(100%LEL)	0 %LEL	----
6/14/2018 11:49:24 AM	O2(40.0%)	20.9 %	----
6/14/2018 11:49:24 AM	H2S(200.0ppm)	0.0 ppm	----
6/14/2018 11:49:24 AM	CO(2000ppm)	0 ppm	----
	---(--)	----	----
6/1/2018 2:29:44 PM	CH4(100%LEL)	0 %LEL	----
6/1/2018 2:29:44 PM	O2(40.0%)	20.9 %	----
6/1/2018 2:29:44 PM	H2S(200.0ppm)	0.0 ppm	----
6/1/2018 2:29:44 PM	CO(2000ppm)	0 ppm	----
	---(--)	----	----
...	Total	3	Datas

- DateTime: Date and time of event occurrence
- Gas: Gas
- Before: Concentration before calibration
- After: Concentration after calibration

Trouble events

DateTime	Gas/Body	Event
6/14/2018 11:49:24 AM	CO(2000ppm)	Fail(Span)
6/14/2018 11:49:24 AM	H2S(200.0ppm)	Fail(Span)
6/14/2018 11:49:24 AM	O2(40.0%)	Fail(Span)
6/14/2018 11:49:24 AM	CH3OH(100%LE	Fail(Span)
6/1/2018 2:29:44 PM	CO(2000ppm)	Fail(Span)
6/1/2018 2:29:44 PM	H2S(200.0ppm)	Fail(Span)
6/1/2018 2:29:44 PM	O2(40.0%)	Fail(Span)
6/1/2018 2:29:44 PM	CH4(100%LEL)	Fail(Span)
...	13 Datas	

DateTime: Date and time of event occurrence
 Gas/Body: Gas generated or GX-3R main unit
 Event: Event type

Bump tests

DateTime	Gas	Test Result	Concentration	Judge
6/14/2018 11:47:0	CH4(100%L	0 %LEL	50 %LEL	FAIL
6/14/2018 11:47:0	O2(40.0%)	21.1 %	12.0 %	FAIL
6/14/2018 11:47:0	H2S(200.0p	0.0 ppm	25.0 ppm	FAIL
6/14/2018 11:47:0	CO(2000ppi	0 ppm	50 ppm	FAIL
	---(--)	----	----	
6/1/2018 2:27:40 F	CH4(100%L	0 %LEL	50 %LEL	FAIL
6/1/2018 2:27:40 F	O2(40.0%)	20.9 %	12.0 %	FAIL
6/1/2018 2:27:40 F	H2S(200.0p	0.0 ppm	25.0 ppm	FAIL
6/1/2018 2:27:40 F	CO(2000ppi	0 ppm	50 ppm	FAIL
	---(--)	----	----	
...	Total	3	Datas	

DateTime: Date and time of event occurrence
 Gas: Gas
 Test Result: Test result concentration
 Concentration: Calibration gas concentration
 Judge: Test assessment

3-4. Data View screen

This screen displays data details in table and graph format.

① Select table or graph. ② Send to printer. ③ Save to file. ④ To view data summary at the same time


No	Date/Time	H4(100%LEL)	O2(40.0%)	I2S(200.0ppm)	CO(2000ppm)	---	temperatur
1	6/15/2018 2:00:08 PM	ALARM H	----	----	----	----	----
2	6/15/2018 2:00:08 PM	ALARM	----	----	----	----	----
3	6/15/2018 2:00:08 PM	WARNING	----	----	----	----	----
4	6/15/2018 2:00:16 PM	AIR	----	AIR	AIR	----	----
5	6/15/2018 2:00:17 PM	29 %LEL	*****	0.0 ppm	0 ppm	----	29.1 °C
6	6/15/2018 2:00:18 PM	-ALARM H	----	----	----	----	----
7	6/15/2018 2:00:18 PM	-ALARM	----	----	----	----	----
8	6/15/2018 2:00:18 PM	-WARNING	----	----	----	----	----
9	6/15/2018 2:00:18 PM	NORMAL	----	----	----	----	----
10	6/15/2018 2:00:27 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.2 °C
11	6/15/2018 2:00:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.3 °C
12	6/15/2018 2:00:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.3 °C
13	6/15/2018 2:00:57 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.4 °C
14	6/15/2018 2:01:07 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.4 °C
15	6/15/2018 2:01:17 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.5 °C
16	6/15/2018 2:01:27 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.6 °C
17	6/15/2018 2:01:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.6 °C
18	6/15/2018 2:01:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.6 °C
19	6/15/2018 2:01:57 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.7 °C
20	6/15/2018 2:02:07 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.7 °C
21	6/15/2018 2:02:17 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.7 °C
22	6/15/2018 2:02:27 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.8 °C
23	6/15/2018 2:02:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.8 °C
24	6/15/2018 2:02:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.8 °C
25	6/15/2018 2:02:57 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.8 °C
26	6/15/2018 2:03:07 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.8 °C
27	6/15/2018 2:03:17 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.9 °C
28	6/15/2018 2:03:27 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.9 °C
29	6/15/2018 2:03:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.9 °C
30	6/15/2018 2:03:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	----	29.9 °C

- Event Only: Displays event data only.
- Condensed: Displays only fluctuating sample data.

CAUTION: No graph will be drawn unless there are at least five samples.

The “Alarm Trend” data table highlights the locations of the alarms occurred in red.

The mouse cursor will appear as follows when hovered over “WARNING”, “ALARM”, or “OVER” event data. Click on the cell here to search for corresponding trend data and to display the data (if any) in a separate window.




Separate window for WARNING, ALARM, and OVER
Click the “Return” button to exit the window.

* The “Alarm Trend” data table highlights the locations of the alarms occurred in red.


① Selecting table or graph

● Select graph

1. Select “Graph” of the “Table” and “Graph” radio buttons at the top left of the screen.



Various operations are available using the checkboxes and combo boxes at the bottom of the screen.



- Upper checkboxes (gas names): Displays or hides the corresponding gas data.
- “Zoom” combo box: Specify a horizontal axis scale factor appropriate for the number of samples.
- “Event” checkbox: Displays event information markers for alarms and other events.
- “Cursor” checkbox: Displays the cursor on the graph.

CAUTION:

The maximum value on the vertical axis of the graph is adjusted automatically based on the following formula:

If the maximum value for data with no events is “x” and the full scale is 10 or greater:

$$Y_{max} = \{\text{int}(x / 10) + 1\} \times 10, \text{ and for full scale under } 10: Y_{max} = \{\text{int}(x) + 1\}$$

“int”: Decimal values are discarded.

CAUTION:

No graph will be drawn unless there are five or more normal concentration data samples.


Data consisting only of events cannot be displayed in graph form because they do not contain concentration information.

The mouse cursor will appear as follows when hovered over “WARNING”, “ALARM”, or “OVER” event data. Click here to search for the corresponding trend data and to display the data (if any) in a separate window.



For WARNING, ALARM, and OVER

Separate window for WARNING, ALARM, and OVER
Click the “Return” button to exit the window.



② Sending to printer

● Print


Details currently displayed on the Data View screen can be sent to the printer.

1. Click "Print" at the top of the screen.

Click "Print".



The printer selection window appears. Select the desired printer and click the "OK" button.




Click the "OK" button to begin printing.

Click the "Cancel" button to return to the Data View screen without printing.

Sample printout (graph print)

Sample printout (table print)



GX-3R Data Logger (Interval Trend) 8/18/2018 10:32:08 AM

Property Value

Name: N15140001

Sampling Date/Time: 8/18/2018 2:00:07 PM to 8/18/2018 2:04:49 PM

Serial No.:

Station ID: S_ID_001

User ID: U_ID_001

Data Count: 28

Interval Time (sec): 10

Gas(Full Scale)	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)	Temperature
Avg	1 %LEL	---	0.0 ppm	0 ppm	---
Max	51 %LEL	---	0.0 ppm	0 ppm	---
Max Date/Time	08/18 14:00:07	---	08/18 14:00:07	08/18 14:00:07	---
Min	---	---	---	---	---
Min Date/Time	---	---	---	---	---
Warning	10 %LEL	18.0 %	1.0 ppm	25 ppm	---
Alarm	50 %LEL	18.0 %	10.0 ppm	50 ppm	---
AlarmH	50 %LEL	20.0 %	10.0 ppm	50 ppm	---
STEL	---	---	5.0 ppm	200 ppm	---
TWA	---	---	1.0 ppm	25 ppm	---

No	Date/Time	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)	Temperature
1	8/18/2018 2:00:08 PM	ALARM H	---	---	---	---
2	8/18/2018 2:00:08 PM	ALARM H	---	---	---	---
3	8/18/2018 2:00:08 PM	WARNING	---	---	---	---
4	8/18/2018 2:00:16 PM	AIR	---	AIR	AIR	---
5	8/18/2018 2:00:17 PM	29 %LEL	---	0.0 ppm	0 ppm	29.4 °C
6	8/18/2018 2:00:18 PM	ALARM H	---	---	---	---
7	8/18/2018 2:00:18 PM	ALARM H	---	---	---	---
8	8/18/2018 2:00:18 PM	WARNING	---	---	---	---
9	8/18/2018 2:00:18 PM	NORMAL	---	---	---	---
10	8/18/2018 2:00:27 PM	0 %LEL	---	0.0 ppm	0 ppm	29.3 °C
11	8/18/2018 2:00:37 PM	0 %LEL	---	0.0 ppm	0 ppm	29.3 °C
12	8/18/2018 2:00:47 PM	0 %LEL	---	0.0 ppm	0 ppm	29.4 °C
13	8/18/2018 2:00:57 PM	0 %LEL	---	0.0 ppm	0 ppm	29.4 °C
14	8/18/2018 2:01:07 PM	0 %LEL	---	0.0 ppm	0 ppm	29.4 °C
15	8/18/2018 2:01:17 PM	0 %LEL	---	0.0 ppm	0 ppm	29.5 °C
16	8/18/2018 2:01:27 PM	0 %LEL	---	0.0 ppm	0 ppm	29.6 °C
17	8/18/2018 2:01:37 PM	0 %LEL	---	0.0 ppm	0 ppm	29.6 °C
18	8/18/2018 2:01:47 PM	0 %LEL	---	0.0 ppm	0 ppm	29.6 °C
19	8/18/2018 2:01:57 PM	0 %LEL	---	0.0 ppm	0 ppm	29.7 °C
20	8/18/2018 2:02:07 PM	0 %LEL	---	0.0 ppm	0 ppm	29.7 °C
21	8/18/2018 2:02:17 PM	0 %LEL	---	0.0 ppm	0 ppm	29.7 °C
22	8/18/2018 2:02:27 PM	0 %LEL	---	0.0 ppm	0 ppm	29.8 °C
23	8/18/2018 2:02:37 PM	0 %LEL	---	0.0 ppm	0 ppm	29.8 °C
24	8/18/2018 2:02:47 PM	0 %LEL	---	0.0 ppm	0 ppm	29.8 °C
25	8/18/2018 2:02:57 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
26	8/18/2018 2:03:07 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
27	8/18/2018 2:03:17 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
28	8/18/2018 2:03:27 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
29	8/18/2018 2:03:37 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
30	8/18/2018 2:03:47 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
31	8/18/2018 2:03:57 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
32	8/18/2018 2:04:07 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
33	8/18/2018 2:04:17 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
34	8/18/2018 2:04:27 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
35	8/18/2018 2:04:37 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C
36	8/18/2018 2:04:47 PM	0 %LEL	---	0.0 ppm	0 ppm	29.9 °C

Sample printout (calibration history)

Sample printout (alarm events)

GX-3R Data Logger (Calibration History) 8/18/2018 10:32:24 AM

Property Value

Serial No.:

Station ID:

User ID:

Last Downloaded: 8/18/2018 10:19:56 AM

No	Date/Time	CH1	CH2	CH3	CH4	CH5
1	8/14/2018 11:49:24 AM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Before	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
	After	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
2	8/1/2018 2:29:44 PM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Before	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
	After	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
3	8/1/2018 2:12:27 PM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Before	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
	After	---	0 %LEL	20.9 %	0.0 ppm	0 ppm

GX-3R Data Logger (Alarm Event) 8/18/2018 10:32:49 AM

Property Value

Serial No.:

Station ID:

User ID:

Last Downloaded: 8/18/2018 10:19:56 AM

No	Date/Time	Gas	Event
1	8/20/2018 4:17:23 PM	CH4(100%LEL)	WARNING
2	8/20/2018 4:17:23 PM	CH4(100%LEL)	ALARM
3	8/20/2018 4:17:23 PM	CH4(100%LEL)	ALARM H

Sample printout (bump tests)

GX-3R Data Logger (Bump Test) 8/18/2018 10:32:38 AM

Property Value

Serial No.:

Station ID:

User ID:

Last Downloaded: 8/18/2018 10:19:56 AM

No	Date/Time	CH1	CH2	CH3	CH4	CH5
1	8/14/2018 11:47:06 AM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Test Result	---	0 %LEL	21.1 %	0.0 ppm	0 ppm
	Concentration	---	50 %LEL	19.0 %	25.0 ppm	50 ppm
	Judge	---	FAIL	FAIL	FAIL	FAIL
2	8/1/2018 2:27:40 PM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Test Result	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
	Concentration	---	50 %LEL	19.0 %	25.0 ppm	50 ppm
	Judge	---	FAIL	FAIL	FAIL	FAIL
3	8/1/2018 2:11:31 PM	Gas	CH4(100%LEL)	O2(40.0%)	H2S(200ppm)	CO(200ppm)
	Test Result	---	0 %LEL	20.9 %	0.0 ppm	0 ppm
	Concentration	---	50 %LEL	19.0 %	25.0 ppm	50 ppm
	Judge	---	FAIL	FAIL	FAIL	FAIL

Sample printout (trouble events)																																																									
<p>GX-3R Data Logger (Trouble Event) 6/18/2018 10:34:05 AM</p> <p>Property Value</p> <p>Serial No. ----- Station ID ----- User ID ----- Last Download 6/18/2018 10:19:56 AM</p> <table border="1"> <thead> <tr> <th>No</th> <th>Date/Time</th> <th>Gas/DO/DOy</th> <th>Event</th> </tr> </thead> <tbody> <tr><td>1</td><td>6/14/2018 11:49:24 AM</td><td>AMCO(2000ppm)</td><td>Fail(Scan)</td></tr> <tr><td>2</td><td>6/14/2018 11:49:24 AM</td><td>H2S(200.0ppm)</td><td>Fail(Scan)</td></tr> <tr><td>3</td><td>6/14/2018 11:49:24 AM</td><td>CO2(40.0%)</td><td>Fail(Scan)</td></tr> <tr><td>4</td><td>6/14/2018 11:49:24 AM</td><td>CO(100%L)</td><td>Fail(Scan)</td></tr> <tr><td>5</td><td>6/11/2018 2:29:44 PM</td><td>CO(2000ppm)</td><td>Fail(Scan)</td></tr> <tr><td>6</td><td>6/11/2018 2:29:44 PM</td><td>H2S(200.0ppm)</td><td>Fail(Scan)</td></tr> <tr><td>7</td><td>6/11/2018 2:29:44 PM</td><td>CO2(40.0%)</td><td>Fail(Scan)</td></tr> <tr><td>8</td><td>6/11/2018 2:29:44 PM</td><td>CO(100%L)</td><td>Fail(Scan)</td></tr> <tr><td>9</td><td>6/11/2018 2:13:27 PM</td><td>CO(2000ppm)</td><td>Fail(Scan)</td></tr> <tr><td>10</td><td>6/11/2018 2:13:27 PM</td><td>H2S(200.0ppm)</td><td>Fail(Scan)</td></tr> <tr><td>11</td><td>6/11/2018 2:13:27 PM</td><td>CO2(40.0%)</td><td>Fail(Scan)</td></tr> <tr><td>12</td><td>6/11/2018 2:13:27 PM</td><td>CO(100%L)</td><td>Fail(Scan)</td></tr> <tr><td>13</td><td>5/20/2018 4:17:19 PM</td><td>CO2(40.0%)</td><td>Fail(Scan)</td></tr> </tbody> </table>	No	Date/Time	Gas/DO/DOy	Event	1	6/14/2018 11:49:24 AM	AMCO(2000ppm)	Fail(Scan)	2	6/14/2018 11:49:24 AM	H2S(200.0ppm)	Fail(Scan)	3	6/14/2018 11:49:24 AM	CO2(40.0%)	Fail(Scan)	4	6/14/2018 11:49:24 AM	CO(100%L)	Fail(Scan)	5	6/11/2018 2:29:44 PM	CO(2000ppm)	Fail(Scan)	6	6/11/2018 2:29:44 PM	H2S(200.0ppm)	Fail(Scan)	7	6/11/2018 2:29:44 PM	CO2(40.0%)	Fail(Scan)	8	6/11/2018 2:29:44 PM	CO(100%L)	Fail(Scan)	9	6/11/2018 2:13:27 PM	CO(2000ppm)	Fail(Scan)	10	6/11/2018 2:13:27 PM	H2S(200.0ppm)	Fail(Scan)	11	6/11/2018 2:13:27 PM	CO2(40.0%)	Fail(Scan)	12	6/11/2018 2:13:27 PM	CO(100%L)	Fail(Scan)	13	5/20/2018 4:17:19 PM	CO2(40.0%)	Fail(Scan)	
No	Date/Time	Gas/DO/DOy	Event																																																						
1	6/14/2018 11:49:24 AM	AMCO(2000ppm)	Fail(Scan)																																																						
2	6/14/2018 11:49:24 AM	H2S(200.0ppm)	Fail(Scan)																																																						
3	6/14/2018 11:49:24 AM	CO2(40.0%)	Fail(Scan)																																																						
4	6/14/2018 11:49:24 AM	CO(100%L)	Fail(Scan)																																																						
5	6/11/2018 2:29:44 PM	CO(2000ppm)	Fail(Scan)																																																						
6	6/11/2018 2:29:44 PM	H2S(200.0ppm)	Fail(Scan)																																																						
7	6/11/2018 2:29:44 PM	CO2(40.0%)	Fail(Scan)																																																						
8	6/11/2018 2:29:44 PM	CO(100%L)	Fail(Scan)																																																						
9	6/11/2018 2:13:27 PM	CO(2000ppm)	Fail(Scan)																																																						
10	6/11/2018 2:13:27 PM	H2S(200.0ppm)	Fail(Scan)																																																						
11	6/11/2018 2:13:27 PM	CO2(40.0%)	Fail(Scan)																																																						
12	6/11/2018 2:13:27 PM	CO(100%L)	Fail(Scan)																																																						
13	5/20/2018 4:17:19 PM	CO2(40.0%)	Fail(Scan)																																																						


Printer setup precautions

- ① The detailed printer settings vary depending on the printer used. Refer to the printer instruction manual for more information.
 - ② This program does not allow the print area to be specified when printing. This means it is not possible to select and print only a certain part of the data view.
 - ③ The setting for the number of copies can be edited only on printers that allow this. Changes in the settings made here will also apply to other applications subsequently used. (For example, if two copies were set here, two copies may also be printed out when using other applications.)
- When printing from other applications after changing the printer settings for this program, check the print settings for that application before printing.

③ Saving to file

● Save

1. Click the “Export” button at the top of the screen.



Specify the destination and file name, then click the “Save” button to save the data. Click the “Cancel” button to cancel saving.

CAUTION:

If a table is displayed, the table contents will be saved in Excel CSV format.
If a graph is displayed, the graph will be saved as a bitmap.

④ Viewing data summary at the same time

● Summary display

1. Click the "Summary" button at the top of the screen.

The screenshot shows the 'Data View(Interval Trend)' window in the GX-3R Data Logger Management Program. The 'Summary' button is highlighted with a callout box labeled 'Click'. Below the property list, a table displays data with a callout box labeled 'A summary is displayed.' The table has columns for No, Date/Time, and various sensor readings.


No	Date/Time	i2S(200.0ppm)	CO(2000ppm)	---	emperatur		
1	6/15/2018 2:00:07 PM	---	---	---	---		
2	6/15/2018 2:00:17 PM	---	---	---	---		
3	6/15/2018 2:00:27 PM	---	---	---	---		
4	6/15/2018 2:00:37 PM	AIR	AIR	---	---		
5	6/15/2018 2:00:47 PM	29 %LEL	*****	0.0 ppm	0 ppm	---	29.1 °C
6	6/15/2018 2:00:57 PM	-ALARM H	---	---	---	---	---
7	6/15/2018 2:01:07 PM	-ALARM	---	---	---	---	---
8	6/15/2018 2:01:17 PM	-WARNING	---	---	---	---	---
9	6/15/2018 2:01:27 PM	NORMAL	---	---	---	---	---
10	6/15/2018 2:01:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.2 °C
11	6/15/2018 2:01:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.3 °C
12	6/15/2018 2:01:57 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.3 °C
13	6/15/2018 2:02:07 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.4 °C
14	6/15/2018 2:02:17 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.4 °C
15	6/15/2018 2:02:27 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.5 °C
16	6/15/2018 2:02:37 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.6 °C
17	6/15/2018 2:02:47 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.6 °C
18	6/15/2018 2:02:57 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.6 °C
19	6/15/2018 2:03:07 PM	0 %LEL	*****	0.0 ppm	0 ppm	---	29.7 °C

Click the "Summary" button while the summary is displayed to hide the summary display.

⑤ Table details

● Event colors

The concentration display cells for each gas in the table have different colored backgrounds based on the type of event that occurred.




Fail:	Gray	Trouble
Warning:	Orange	1st alarm
Alarm:	Red:	2nd alarm
AlarmH:	Magenta	3rd alarm
TWA:	Bright purple	TWA alarm
STEL:	Pink	STEL alarm
Normal:	Dark green	Restored from above statuses
Over:	Bright red	Full-scale over

⑥ Graph details

● Cursor

1. Select the “Cursor” checkbox to display the cursor over the graph.




Use the “←” and “→” keys to move the cursor left and right. Use the “↑” and “↓” keys to move the time and concentration display up and down. Press the “Shift” key at the same time to move the cursor faster.

CAUTION:

The cursor cannot be moved if a window for another program is opened and the focus is not currently on the graph area.

In this case, the cursor appears as a broken line. To return the focus, click anywhere within the graph area.



3-5. Last Calibration screen

This checks for the calibration expiration of previously downloaded main unit data. Bump tests are also displayed in the same way.

The screenshot shows the GX-3R Data Logger software interface. It features two main data tables: 'Last Calibration' and 'Last Bump Test'. The 'Last Calibration' table has columns for No., SerialNo, UserID, StationID, Gas, Before, After, A.Cal., and Cal.Due(Day). The 'Last Bump Test' table has columns for No., SerialNo, UserID, StationID, Gas, Test Result, Concentration, and Bump Test Du. A right-hand sidebar contains several buttons: Download, Instrument Information, Data, Last Calibration, Set, and Exit. Callouts point to specific elements: 'Select display details.' points to the 'Calibration Record' radio button; 'Send to printer.' points to the 'Print' button; 'Delete data.' and 'Change password.' point to a greyed-out area below the calibration table; 'Click this button.' points to the 'Last Calibration' sidebar button; and 'Bump test information' points to the 'Last Bump Test' table.

No.	SerialNo	UserID	StationID	Gas	Before	After	A.Cal.	Cal.Due(Day)
1		U_ID_001	S_ID_001	CH4	0	0	50	Now
				O2	0.0	0.0	12.0	Now
				H2S	0.0	0.0	25.0	Now
				CO	0	0	50	Now

No.	SerialNo	UserID	StationID	Gas	Test Result	Concentration	Bump Test Du
1		U_ID_001	S_ID_001	CH4	0	0	Now
				O2	0.0	0.0	Now
				H2S	0.0	0.0	Now
					0	0	Now

CAUTION: The table details are read-only and cannot be edited.

① Selecting display details

● Expired data

1. Click the “Need Calibration” radio button.

Last Calibration									
<input checked="" type="radio"/> Need Calibration									
<input type="radio"/> Calibration Date									
<input type="radio"/> Calibration Record									
Print									
No.	SerialNo	UserID	StationID	CH4	O2	H2S	CO		Last Downl
1	-----	U_ID_001	S_ID_001		1/18	1/1/2018	1/1/2018	----	6/18/2018 1

Displays calibrations for previously connected GX-3R main units (for which device information data has been downloaded) that have expired.

● List display

1. Click the “Calibration Date” radio button.

Last Calibration									
<input type="radio"/> Need Calibration									
<input checked="" type="radio"/> Calibration Date									
<input type="radio"/> Calibration Record									
Print									
No.	SerialNo	UserID	StationID	CH4	O2	H2S	CO		Last Downl
1	-----	U_ID_001	S_ID_001	1/1/2018	1/1/2018	1/1/2018	1/1/2018	----	6/18/2018 1

Lists data for previously connected GX-3R main units. (Only the most recent data is listed for units with the same serial number, user ID and station ID.)

● Detailed display

1. Click the “Calibration Record” radio button.

Last Calibration									
<input type="radio"/> Need Calibration									
<input type="radio"/> Calibration Date									
<input checked="" type="radio"/> Calibration Record									
Print									
No.	SerialNo	UserID	StationID	Gas	Before	After	A.Cal.	Cal.Due(Day)	
1	-----	U_ID_001	S_ID_001	CH4		0	0	50	Now
				O2		0.0	0.0	12.0	Now
				H2S		0.0	0.0	25.0	Now
				CO		0	0	50	Now
					----	----	----	----	


Displays data for previously connected GX-3R main units in the same format as the Instrument Information screen.

For more information on display contents: → Refer to ③ Calibration history information in 3-2. Instrument Information screen.

② Sending to printer

● Print

The most recent calibration dates shown in the “Need Calibration” and “Calibration Date” displays can be printed out.
The printer selection window appears. Select the desired printer and click the “OK” button.




GX-3R Data Logger (Last Calibration)										6/18/2018 10:39:52 AM
No.	Serial No	UserID	StationID	CH4	CO2	H2S	CO	----	Last Download	
1		S_ID_001	S_ID_001	1/1/2018	1/1/2018	1/1/2018	1/1/2018	----	6/18/2018 10:16:26 AM	

③ Deleting data

● Delete

1. Right-click over the data to delete.



CAUTION: Only data shown in the “Need Calibration” and “Calibration Date” displays can be deleted. Data cannot be deleted in the “Calibration Record” display.


● Password input

1. Click the “Delete” button to display a password dialog. Enter the password, then click the “Continue” button.



CAUTION: Clicking “Continue” without entering a password will cancel the delete.

2. The following message will appear when you enter the correct password and click the “Continue” button.



Click the “Yes” button to delete the data.
Click the “No” button to cancel data deletion.

④ Changing the password

● Password input

1. Open the password dialog in the same way as for data deletion, then click the “Change Password” button.



2. Enter the correct password, then click the “Current password” button. The following message appears.



3. Enter the new password here, then click the “New Password” button.

4. A password dialog will appear again. Enter the same (new) password, then click the “New Password” button.




Lastly, click “OK” to update to the new password.

CAUTION: The default password immediately after installation is "riken" (not case-sensitive).

3-6. Set screen

This screen is used to configure screen display settings and main unit status settings.



CAUTION: Data that has been set or changed must be sent to the GX-3R main unit by clicking the “Update” button.

CAUTION: Font settings will be applied from the next time the program is started.


CAUTION: “BLE auto shutoff when idle” and “BLE Program No” are displayed only when connected with GX-3R Pro with BLE function.

CAUTION: “Automatic start after successful bump test” and “Automatic start after successful calibration” may not be displayed depending on the software version.

① Setting font and graph colors

● Change font

1. Click the font display area.



Set the desired font in the font setting dialog that appears.


CAUTION:

The screen display may become hard to read if you use an excessively large font. Changes made here will be applied from the next time the program is started.

● Change graph color

The display colors for respective gases can be changed on the graph.


1. Click the color area for a gas.



Select the desired color in the color selection dialog that appears.

● Edit station/user lists

1. The station ID and user ID lists can be edited. Click "Detail Settings".



An edit dialog appears. The station list can contain up to 128 entries. You can use up to 16 alphanumerical characters, including the space, hyphen, and slash symbols.

Export csv file : Outputs to file. Creates a text file containing the number and data entries.

Import csv file : Imports from a text file containing number and data entries. The background color will appear in red for entries with unsupported characters or too many characters.

* **OK** cannot be clicked if there are entries with a red background.



② Changing main unit status

● Edit

2. Edit data in the status area as required.

Details indicated for “Serial No.”, “Station ID”, and “User ID” can be edited provided they are within 16 characters.

“Interval Trend Time” can be changed by selecting from the pull-down list.

Click the “Date/Time Set” button to synchronize the PC time (“PC Date/Time”) with the GX-3R main unit internal clock (“GX-3R Date/Time”).

When “BLE auto shutoff when idle” checkbox is selected, BLE function turns off in 5 minutes of inactivity after turning ON.

When “Automatic start after successful bump test” checkbox is selected, GX-3R/GX-3R Pro will automatically start the startup procedure after bump test.

When “Automatic start after successful calibration” checkbox is selected, GX-3R/GX-3R Pro will automatically start the startup procedure after calibration.

CAUTION: Dates and times cannot be entered directly into the date/time boxes.

CAUTION: Apart from clock setting, changing the status data here alone will not update the same data in the GX-3R main unit. Be sure to click “Update” to update the changes in the main unit.


CAUTION: “BLE auto shutoff when idle” and “BLE Program No” are displayed only when connected with GX-3R Pro with BLE function.

CAUTION: “Automatic start after successful bump test” and “Automatic start after successful calibration” may not be displayed depending on the software version.

③ Updating changes in GX-3R main unit

● Notify of changes

1. Click the "Update" button after making changes.



Click the "Yes" button to send the changes to the GX-3R main unit to be stored.
Click the "No" button to cancel update notification.

CAUTION: You cannot restore changed details. If "Update" has not yet been clicked, you can restore the changed details to the information on the main unit by clicking the "Instrument Information" button on the "Download" screen to download the device information data.

4. Data Maintenance

Depending on how the program is used, if data is read in several times a day, the growing volumes of data may make it harder to find specific data. Unforeseen problems with the PC may also lead to loss of valuable data.

We recommend backing up data periodically to protect against such circumstances.

4-1. Data storage configuration details

Data is contained within the installed GX-3R program folder.

- 1) File name: GX3R.mdb
File type: Microsoft JET 3.6 database file

- 2) File name: Data
File type: Folder. Individual trend data files are found within the corresponding year and month folders.

4-2. Backing up

Depending on the system configuration, we recommend copying data to a separate hard disk drive or external auxiliary storage device (for example, MO or CD-R).

When restoring data, copy the data to the location at which the GX-3R executable file is saved. This will allow data to be searched and viewed when the program is started.

5. Usage Precautions

Note the following precautions when using this program:

- ① Confirm that the GX-3R is in a suitable location for receiving data. Normal communication is not possible if it is not in a suitable location.
- ② Avoid using similar functions on other applications at the same time when data is being received. (For example, using infrared communication with other applications while data is being received)
- ③ Do not forcibly shut down this program. (For example, using the Ctrl + Alt + Del operation) The program performs shutdown processing in which it saves configuration parameters for use the next time the program starts. Forcibly shutting down the program may cause problems the next time the program starts.
- ④ Do not directly modify or overwrite data files.

6. Troubleshooting

Symptoms	Cause	Corrective action
Communication is not possible.	The main unit is not in a suitable location. Obstacles are in the way.	Reposition the main unit.
	Other devices using infrared are present.	Turn off the other devices or take other measures to prevent interference.
An error occurs during communication.	Disturbance light is present.	Remove any other devices that use infrared.
	The GX-3R main unit moves during communication.	Make sure the GX-3R main unit does not move during communication.
The communication data is erroneous.	Disturbance light is present.	Remove any other devices that use or emit infrared.

If the problems persist even after taking the action described here, please contact Riken Keiki.

7. IrDA Specifications

7-1. Infrared communication

Infrared communication (IrDA protocol) is used for communicating with the main unit.

Check whether the PC being used supports infrared communication.

Ensure a direct line of sight between the PC communication port and the infrared communication port on the main unit. Remove any other sources of light interference.

CAUTION:

This program is able to communicate with the GX-3R main unit only in environments that support IrDA. Confirm that the PC being used has a built-in IrDA device that is currently usable.

An IrDA USB adapter is required if you are using a PC that lacks a built-in IrDA device (for example, most desktop PCs and some laptop PCs).

8. File Organization

Details of the files present when the program is installed and the files present during operation are provided below.

8-1. Current directory immediately after installation

File name	Details
GX3R.exe RklrDA11.ocx Filemove.avi	GX-3R program body Infrared communication component Animation file used during data reception

8-2. Current directory during operation

File name	Details
GX3R.exe RklrDA11.ocx Filemove.avi	GX-3R program body Infrared communication component Animation file used during data reception
GX3R.ini GX3R.dat GX3R.mdb Data Serial.log	GX-3R initial setup file Data downloading file Database file (Microsoft Jet 3.6 database) Directory for saving trend data files Record of communication port details since the program was started (for investigation/maintenance use)

CAUTION: The files and directories listed below the double line are created after the program is started.

9. Software Function Specifications

Product name (Program name)	GX-3R Data Logger Management Program
Product model	
Executable file name	GX3R.EXE
Compatible operating systems	Microsoft Corporation Windows 7 Windows 8 Windows 10
Program size	Main program approx. 3 MB; libraries approx. 5.2 MB (Uses up to 40 MB of hard disk space during installation.)
Communication with main unit	Infrared (IrDA 1.1 protocol) compliant method Standard communication settings Baud rate 57,600 bps (upper limit) Data bits 8 bits Stop bit 1 bit Parity Even
Transmission time	Max. approx. 3 minutes (for maximum number of data samples and standard communication settings)
Medium	CD-ROM ×1
Package contents	Operating manual (this document) Product warranty registration card User license agreement

Manual Log

Rev.	Amendment	Issue data
0	First issue	2019/4/1
1	P.40 Initial password correction	2019/6/6
2	P40/43 Function added	2019/12/12