

EC-TYPE EXAMINATION CERTIFICATE (MODULE B)

Certificate No:
MEDB000074D
Revision No:
3

Application of: Directive 2014/90/EU of 23 July 2014 on marine equipment (MED). This Certificate is issued by DNV SE based on the notification of the Federal Maritime and Hydrographic Agency of Germany.

This is to certify:

That the Portable Oxygen analysis and gas detection equipment: category 2: (explosive gas atmospheres)

with type designation(s)

Portable Gas Monitor GX-3R & GX-3R Pro

Issued to

RIKEN KEIKI Co., Ltd.
Itabashi-ku, Tokyo, Japan

is found to comply with the requirements in the following Regulations/Standards:

Regulation (EU) 2024/1975,

item No. MED/3.30b. SOLAS 74, Regulation II-2/4, VI/3, XI-1/7, FSS Code 15 and MSC.1/Circ.1477, IMO MSC.1/Circ.1581.


Further details of the equipment and conditions for certification are given overleaf.

This Certificate is valid until **2030-02-04**.

Issued at **Hamburg** on **2025-02-05**

DNV local station:
Japan CMC

Approval Engineer:
Dariusz Lesniewski



Notified Body
No.: **0098**



for **DNV GL SE**

Digitally Signed By:
Christine Mydlak-Röder
Location: DNV Hamburg,
Germany

Mydlak-Röder, Christine
Head of Notified Body

The mark of conformity may only be affixed to the above type approved equipment and a Manufacturer's Declaration of Conformity issued when the production-surveillance module (D, E or F) of Annex B of the MED is fully complied with and controlled by a written inspection agreement with a Notified Body. The product liability rests with the manufacturer or his representative in accordance with Directive 2014/90/EU.

This certificate is valid for equipment, which is conform to the approved type. The manufacturer shall inform DNV SE of any changes to the approved equipment. This certificate remains valid unless suspended, withdrawn, recalled or cancelled.

Should the specified regulations or standards be amended during the validity of this certificate, the product is to be re-approved before being placed on board a vessel to which the amended regulations or standards apply.

LEGAL DISCLAIMER: Unless otherwise stated in the applicable contract with the holder of this document, or following from mandatory law, the liability of DNV AS, its parent companies and their subsidiaries as well as their officers, directors and employees ("DNV") arising from or in connection with the services rendered for the purpose of the issuance of this document or reliance thereon, whether in contract or in tort (including negligence), shall be limited to direct losses and under any circumstance be limited to 300,000 USD.



Product description

Portable Gas Monitor GX-3R/3R Pro is a gas monitor designed to measure the combustible gases (LEL), oxygen (O₂), hydrogen sulfide (H₂S), and carbon monoxide (CO). The GX-3R Pro can additionally measure toxic gases¹⁾ including carbon dioxide and sulfur dioxides (CO₂ and SO₂). The gas sampling is by diffusion. The model does not have internal suction pump. Li-ion battery unit BP-3R (for GX-3R) and BUL-3R (for GX-3R Pro) are for power supply.

[GX-3R]

Detection target gas	HC or CH ₄	O ₂	CO	H ₂ S
Detection principle	New ceramic type	Electrochemical type	Electrochemical type	Electrochemical type
Detection range <Service range>	0 - 100.0 %LEL	0 - 25.0 vol% < - 40.0 vol%>	0 - 500 ppm < - 2000 ppm>	0 – 100.0 ppm < - 200.0 ppm>
1 digit	1 %LEL	0.1 vol%	1 ppm	0.1 ppm

Power supply	Lithium ion rechargeable battery [BP-3R]
Dimensions	Approx. 58mm (W) × 65mm (H) × 26mm (D) (excluding protrusions)
Weight	Approx. 100g
Operating Temp.	-20 °C to +50 °C (no sudden changes) (Continuous use environment)
Ingress proof rating	IP68 (2m, 1h)

[GX-3R Pro]

Detection target gas	HC or CH ₄	O ₂	CO	H ₂ S
Detection principle	New ceramic type	Electrochemical type	Electrochemical type	Electrochemical type
Detection range <Service range>	0 - 100.0 %LEL	0 - 25.0 vol% < - 40.0 vol%>	0 - 500 ppm < - 2000 ppm>	0 – 100.0 ppm < - 200.0 ppm>
1 digit	1 %LEL	0.1 vol%	1 ppm	0.1 ppm
Detection target gas	SO ₂ ¹⁾	CO ₂ ¹⁾		
Detection principle	Electrochemical type	Infrared type		
Detection range <Service range>	0 – 20.00 ppm < - 100.00 ppm>	0 - 5.00 vol% < - 10.00 vol%>	0 - 10000 ppm	
1 digit	0.05 ppm	0.01 vol%	20 ppm	

Power supply	Lithium ion rechargeable battery [BUL-3R]
Dimensions	Approx. 73mm (W) × 65mm (H) × 26mm (D) (excluding protrusions)
Weight	Approx. 120g
Operating Temp.	-20 °C to +50 °C (no sudden changes) (Continuous use environment)
Ingress proof rating	IP68 (2m, 1h)

Note: ¹⁾ Toxic gas detection not part of MED/3.30 homologation as documented by present certificate.

Place of manufacture


RIKEN KEIKI Co., Ltd.
 2-3, Minamisakae-cho, Kasukabe-shi,
 Saitama, 344-0057 Japan

RIKEN KEIKI NARA MFG. Co., Ltd.
 49-1 Abe, Sakurai-shi,
 Nara, 633-0054 Japan

Application/Limitation

- Thermal shock test and Oil resistance test in accordance with IEC 60945: 2002 (Ed.4) additionally carried out (see report No. 20-284(E) dated 2020.11.19 and 20-329(E) dated 2020.12.03 and 20-334(E) dated 2020.12.04).
- The equipment is compliant with following DNV location Class and application:
 - EMC-A : All locations except bridge and open deck (IEC 60092-504 (2016) table 1 item 15-20)
 - Enclosure-D : IP68 (Location according ref. IEC 60529:2001)
 - VIB-A : General applications (ref. IEC 60092-504 (2016) table 1 item 10)
 - TEMP-C : Location (-25°C to +45°C) (ref. IEC 60092-504 (2016) table 1 item 6-7),
 - Compliant with IACS UR E10 Rev. 09

- Ex-certification is not covered by this certificate. Application in hazardous area to be approved in each case according to the Rules and Ex-Certification/Special Condition for Safe Use listed in valid Ex-Certificate issued by a notified/recognized Certification Body.
 The following paragraph provided for information only, is based on information received from the manufacturer, but not verified by DNV.

Information on Ex-Certification received from manufacturer – Not verified by DNV		
Equipment	Certified	Certificate No.
[GX-3R] [GX-3R Pro]	 II 1 G Ex da ia IIC T4 Ga or Ex ia IIC T4 Ga and I M1 Ex da ia I Ma or Ex ia I Ma, -40°C to +60°C	DEKRA 17ATEX0103 X

- Each Equipment delivery to be provided with installation and user manual.

Type Examination documentation

Hidden

DEKRA 17ATEX0103 X issued by DEKRA Certification B.V. Issuing date 2020.07.07

Report	Issuing date	Drawings
No. 20-260(E)	2020.11.10	M3-4777-14-01K
No. 20-261(E)	2020.11.10	M3-4777-14-02K
No. 20-281(E)	2020.11.18	M3-4777-15-01K
No. 20-282(E)	2020.11.17	M3-4777-15-02K
No. 20-283(E)	2020.11.17	M3-4777-15-03K
No. 20-284(E)	2020.11.19	M3-4777-15-04K
No. 20-329(E)	2020.12.03	E4-6991-5797-70-01K
No. 20-334(E)	2020.12.04	E4-6991-5798-40-01K
No. 382-180023-1	2018.08.10	
No. 382-180023-2	2018.08.10	
No. 400-2000035	2020.08.31	
No. 400-2000036-1	2020.08.31	
No. 200989-1	2020.10.14	
No. 200989-2	2020.10.14	
No. 41500120P	2020.07.30	

Re-issuing to revision 01

EN 50104:2010 to EN 50104:2019

IEC 60079-0:2018 to EN IEC 60079-0:2018 incl. AC:2020

TEST REPORT No. 41500120P.

page 7 of 8 of 224649000-ExTR17.0063.02-Cover

Re-issuing to revision 03

DEKRA TEST Report No. 41300121P NII

EU-Type Examination Certificate DEKRA 17ATEX0103X Issue No. 5

Declaration of Conformity for GX-3R and GX3R Pro Doc. No. MM-MCC_7290421_DOC_Rev.0

Tests carried out

IEC 60945:2002 incl. IEC 60945 Corr. 1: 2008 or IEC 60945:2002 incl. IEC 60945 Corr. 1:2008,
 IEC 60092-504:2016,
 IEC 60533:2015

Category 2: (explosive gas atmospheres):

- EN 50104:2019 incl. EN 50104:2019/A1:2023,
- EN 60079-29-1:2016 incl. A1:2022 and A11:2022,
- EN IEC 60079-0:2018 incl. AC2020,
- EN 60079-1:2014 incl.AC:2018-09,
- EN IEC 60079-10-1:2021*
- EN 60079-11:2012
- EN IEC 60079-15:2019*
- EN 60079-26:2015*

*Standard not applicable



Job Id: **344.1-011097-4**
Certificate No: **MEDB000074D**
Revision No: **3**

Marking of product

For identification to this type examination certificate the products shall be marked with:

- Manufacturer's name or trademark
- Type designation
- Mark of Conformity (see front page)

END OF CERTIFICATE